

Agilent 4285A

Precision LCR Meter

Data Sheet

Specifications

The complete Agilent Technologies 4285A specifications are listed below. These specifications are the performance standards or limits against which the instrument is tested. When shipped from the factory, the 4285A meets the specifications listed in this section. The specification test procedures are covered in *Agilent 4285A Maintenance Manual* (Agilent Part Number 04285-90030).

Measurement Functions

Measurement parameters

- |Z| = Absolute value of impedance
- |Y| = Absolute value of admittance
- L = Inductance
- C = Capacitance
- R = Resistance
- G = Conductance
- D = Dissipation factor
- Q = Quality factor
- R_s = Equivalent series resistance
- R_p = Parallel resistance
- X = Reactance
- B = Susceptance
- θ = Phase angle

Combinations of measurement parameters

Z , Y	L, C	R	G
θ (deg), θ (rad)	D, Q, R _s , R _p , G	X	B

Mathematical functions

The deviation and the percent of deviation of measurement values from a programmable reference value.

Equivalent measurement circuit

Parallel and series

Ranging

Auto and manual (hold/up/down)

Trigger

Internal, external, BUS (GPIB), and manual

Delay time

Programmable delay from the trigger command to the start of the measurement, 0 to 60.000 s in 1 ms steps.

Measurement terminals

Four-terminal pair

Test cable length

0 m, 1 m, and 2 m selectable

Integration time

Short, medium, and long selectable (refer to *Supplemental Performance Characteristics*, page 14, for the measurement time)

Averaging

1 to 256, programmable



Test Signal

Frequency

75 kHz to 30 MHz, 100 Hz solution

Frequency accuracy

±0.01%

Signal modes

Normal—Program selected voltage or current at the measurement terminals when they are opened or shorted, respectively

Constant—Maintains selected voltage or current at the device under test (DUT) independent of changes in the device's impedance.

Signal level

The following test signal level accuracy is specified for an ambient temperature range of 23 °C ± 5 °C and the test cable length is 0 m.

	Mode	Range	Setting accuracy
Voltage	Normal	5 m V _{rms} to 2 V _{rms}	±{(8 + 0.4 f _m)% + 1 m V _{rms} }
	Constant ¹	10 mV _{rms} to 1 V _{rms}	±{(6 + 0.2 f _m)% + 1 m V _{rms} }
Current	Normal	200 μA _{rms} to 20 mA _{rms}	±{(8 + 1 f _m)% + 40 μA _{rms} }
	Constant ¹	100 μA _{rms} to 20 mA _{rms}	±{(6 + 0.2 f _m)% + 40 μA _{rms} }

1. When the ALC function is set to ON

For the temperature range of 0 °C to 55 °C, multiply the temperature induced setting error listed in Figure 1-5 to the test signal setting accuracy. When test cable length is 1 m or 2 m, add the following error due to test cable length.

$$0.2 \times f_m \times L [\%]$$

where:

f_m = Test frequency [MHz]

L = Test cable length [m]

Output impedance

The following output impedance is specified for the test cable length of 0 m:

$$(25 + 0.5 f_m) \Omega \pm \left(10 + \frac{2}{3} f_m\right) \%$$

where:

f_m = Test frequency [MHz]

Test signal level monitor

The following test signal level monitor accuracy is specified for an ambient temperature range of 23 °C ± 5 °C and the test cable length is 0 m.

Mode	Range	Monitor accuracy
Voltage	0.01 mV _{rms} – 2.000 V _{rms}	$\pm \left\{ S_{mon} \frac{4 + 0.2 f_m}{100} + \frac{S_{set}}{500} \right\} [V]$
Current	0.001 μA _{rms} – 20.00 mA _{rms}	

For the temperature range of 0 °C to 55 °C, multiply the temperature induced setting error listed in Figure 1-5. When test cable length is 1 m or 2 m, add the following error due to test cable length.

$$S_{mon} \times 0.2 \times f_m \times L [V]$$

where:

f_m = Test frequency [MHz]

L = Test cable length [m]

S_{mon} = Readout value of test signal level

S_{set} = Setting value of test signal level

For example,

Test frequency:	1 MHz
Test signal level:	1 V _{rms}
Monitor readout value:	500 mV _{rms}
Cable length:	1 m
Ambient temperature:	25 °C

Then, voltage level monitor accuracy V_{ma} is

$$\begin{aligned}\Delta V_{ma} &= 0.5 \times \frac{4 + 0.2 \times 1}{100} + \frac{1 \times 0.2}{100} + 0.5 \times \frac{0.2}{100} \times 1 \times 1 \\ &= 0.024 \text{ [V]} \\ V_{ma} &= \frac{0.024}{0.5} \\ &\approx 4.8 \text{ [%]}\end{aligned}$$

Display Range

Parameter	Range
Z , R, X	0.00001 Ω to 99.9999 M Ω
Y , G, B	0.00001 μ S to 99.9999 S
C	0.00001 pF to 999.999 μ F
L	0.001 nH to 99.9999 H
D	0.000001 to 9.99999
Q	0.01 to 99999.9
θ	-180.000° to 180.000°
Δ	-999.999% to 999.999%

Measurement Accuracy

The measurement accuracy includes stability, temperature coefficient, linearity, repeatability, and calibration interpolation error. The measurement accuracy is specified when all of the following conditions are satisfied:

- Warm-up time: ≥ 30 minutes
- Test cable length: 0 m, 1 m (Agilent 16048A), or 2 m (Agilent 16048D). For the 1 m or 2 m cable length operation (with Agilent 16048A/D), CABLE CORRECTION has been performed
- OPEN and SHORT corrections have been performed.
- The optimum measurement range is selected by matching the DUT's impedance to the effective measuring range shown in Figure 1-1 and Figure 1-2. (For example, if the DUT's impedance is 3 k Ω and oscillator level is less than or equal to 1 V, the optimum range is the 500 Ω range.)
- Measurement accuracy is specified at the following reference planes:
 - Test frequency ≤ 1 MHz
At the UNKNOWN terminals on the Agilent 4285A front panel or at the end of the standard test leads (Agilent 16048A/D).
 - Test frequency ≥ 1.001 MHz
At the 1-port terminal of the Agilent 16085B Terminal Adapter, which should be connected to the UNKNOWN terminals of the Agilent 4285A or to the end of the standard test leads (Agilent 16048A/D).




Figure 1-1. Effective measurement range (oscillator level $\leq 1 V_{rms}$)




Figure 1-2. Effective measurement range (oscillator level > 1 V_{rms})

|Z|, |Y|, L, C, R, X, G, and B accuracy

|Z|, |Y|, L, C, R, X, G, and B accuracy A_e is given as

$$A_e = \pm(A_n + A_c) \times K_t [\%]$$

where:

A_n = Basic accuracy equation given from the A_1 to A_{16} shown in Table 1-1. The applicable frequency range and impedance range of equations A_1 to A_{16} are shown in Figure 1-3 and Figure 1-4. (Refer to *Basic Accuracy Equations* on page 7.)

A_c = Cable length factor (Refer to *Cable Length Factor* on page 10.)

K_t = Temperature factor (Refer to *Temperature Factor* on page 10.)

L, C, X, and B accuracies apply when D_x (measured D value) ≤ 0.1 .

When $D_x > 0.1$, multiply A_e by $\sqrt{1 + D_x^2}$ for L, C, X, and B accuracies.

R and G accuracies apply when Q_x (measured Q value) ≤ 0.1 .

When $Q_x > 0.1$, multiply A_e by $\sqrt{1 + Q_x^2}$ for R and G accuracies. G accuracy given by the equation above applies to the G-B combination only.

D accuracy

D accuracy D_e is given as

$$D_e = \pm \frac{A_e}{100}$$

where:

A_e = |Z|, |Y|, L, C, R, X, G, and B accuracy

D accuracy applies when D_x (measured D value) ≤ 0.1 .

When $D_x > 0.1$, multiply D_e by $(1 + D_x)$.

Q accuracy

Q accuracy Q_e is given as

$$Q_e = \pm \frac{Q_x \times D_e}{1 \mp Q_x \times D_e}$$

where:

Q_x = Measured Q value

D_e = D accuracy

Q accuracy applies when $Q_x \times D_e < 1$.

 θ Accuracy

θ accuracy θ_e is given as

$$\theta_e = \pm \frac{180 \times A_e}{\pi \times 100} \text{ [deg]}$$

where:

A_e = |Z|, |Y|, L, C, R, X, G, and B accuracy

G Accuracy

G accuracy G_e is given as

$$G_e = \pm B_x \times D_e \quad [S]$$

$$\left(B_x = 2 \pi f C_x = \frac{1}{2 \pi f L_x} \right)$$

where:

B_x = Measured B value [S]

C_x = Measured C value [F]

L_x = Measured L value [H]

D_e = D accuracy

f = Test frequency [Hz]

G accuracy applies when D_x (measured D value) ≤ 0.1 .

G accuracy given by the equation above applies to the C_p -G and L_p -G combinations only.

R_p accuracy

R_p accuracy R_{pe} is given as

$$R_{pe} = \pm \frac{R_{px} \times D_e}{D_x \mp D_e} \quad [\Omega]$$

where:

R_{px} = Measured R_p value [Ω]

D_x = Measured D value

D_e = D accuracy

R_p accuracy applies when D_x (measured D value) ≤ 0.1.

R_s accuracy

R_s accuracy R_{se} is given as

$$R_{se} = \pm X_x \times D_e \quad [\Omega]$$

$$\left(X_x = 2 \pi f L_x = \frac{1}{2 \pi f C_x} \right)$$

where:

X_x = Measured X value [Ω]

C_x = Measured C value [F]

L_x = Measured L value [H]

D_e = D accuracy

f = Test frequency [Hz]

R_s accuracy applies when D_x (measured D value) ≤ 0.1.

Basic accuracy equations

The basic accuracy A_n is calculated from the following procedure:

1. Determine A_n equation from Figure 1-3 or Figure 1-4. In both charts, boundary line belongs to the better accuracy area.

When the oscillator level is ≤ 1 V_{rms} , determine A_n to be applied, value of K_i and value of K_{osc} from the Figure 1-3. If the determined K_{osc} ≤ 1, then round up K_{osc} to 1.

When the oscillator level is > 1 V_{rms} , determine A_n to be applied and value of K_i from the Figure 1-4.

2. Calculate A_n from the formula to be applied. The n accuracy factor included in the A_n equation is shown in Table 1-2. Use K_i and K_{osc} factors determined in previous step.

Table 1-1. A_n equations

$A_1 = N_1\% + \left(\frac{f_m}{30}\right)^2 \cdot 3\% + \frac{50}{ Z_m } [0.02\% + \left(\frac{f_m}{30}\right) \cdot 0.1\%] \cdot K_i \cdot K_{osc}$
$A_2 = N_1\% + \left(\frac{f_m}{30}\right)^2 \cdot 3\% + \frac{ Z_m }{50} [0.02\% + \left(\frac{f_m}{30}\right) \cdot 0.05\%] \cdot K_i \cdot K_{osc}$
$A_3 = N_1\% + \left(\frac{f_m}{5}\right)^2 \cdot 0.1\% + \frac{ Z_m }{500} [0.02\% + \left(\frac{f_m}{30}\right) \cdot 0.05\%] \cdot K_i \cdot K_{osc}$
$A_4 = 0.3\% + \left(\frac{f_m}{30}\right)^2 \cdot 3\% + \frac{ Z_m }{500} [0.05\% + \left(\frac{f_m}{30}\right) \cdot 0.1\%] \cdot K_i \cdot K_{osc}$
$A_5 = 0.18\% + \frac{ Z_m }{5k} \cdot 0.02\% \cdot K_i \cdot K_{osc}$
$A_6 = 0.18\% + \left(\frac{f_m}{30}\right)^2 \cdot 3\% + \frac{ Z_m }{5k} [0.02\% + \left(\frac{f_m}{10}\right) \cdot 0.03\%] \cdot K_i \cdot K_{osc}$
$A_7 = 0.5\% + \left(\frac{f_m}{30}\right)^2 \cdot 3\% + \frac{ Z_m }{5k} \cdot \left(\frac{f_m}{30}\right) \cdot 0.2\% \cdot K_i \cdot K_{osc}$
$A_8 = 0.18\% + \frac{ Z_m }{50k} \cdot 0.03\% \cdot K_i \cdot K_{osc}$
$A_9 = N_2\% + \left(\frac{f_m}{30}\right)^2 \cdot 3\% + \frac{100}{ Z_m } [0.02\% + \left(\frac{f_m}{30}\right) \cdot 0.1\%] \cdot K_i$
$A_{10} = N_2\% + \left(\frac{f_m}{30}\right)^2 \cdot 3\% + \frac{ Z_m }{100} [0.02\% + \left(\frac{f_m}{30}\right) \cdot 0.05\%] \cdot K_i$
$A_{11} = 0.18\% + \left(\frac{f_m}{5}\right)^2 \cdot 0.1\% + \frac{ Z_m }{1k} [0.02\% + \left(\frac{f_m}{30}\right) \cdot 0.05\%] \cdot K_i$
$A_{12} = 0.3\% + \left(\frac{f_m}{30}\right)^2 \cdot 3\% + \frac{ Z_m }{1k} [0.05\% + \left(\frac{f_m}{30}\right) \cdot 0.1\%] \cdot K_i$
$A_{13} = 0.18\% + \frac{ Z_m }{10k} \cdot 0.02\% \cdot K_i$
$A_{14} = 0.18\% + \left(\frac{f_m}{30}\right)^2 \cdot 3\% + \frac{ Z_m }{10k} [0.02\% + \left(\frac{f_m}{10}\right) \cdot 0.03\%] \cdot K_i$
$A_{15} = 0.5\% + \left(\frac{f_m}{30}\right)^2 \cdot 3\% + \frac{ Z_m }{10k} \cdot \left(\frac{f_m}{30}\right) \cdot 0.2\% \cdot K_i$
$A_{16} = 0.18\% + \frac{ Z_m }{100k} \cdot 0.03\% \cdot K_i$

$|Z_m|$: Absolute value of measured impedance in [Ω]

f_m : Test frequency in [MHz]

N accuracy factors

N_1 and N_2 in the A_n equations have the following values:

Table 1-2. N accuracy factors

Frequency(f)	N_1	N_2
75 kHz ≤ f ≤ 200 kHz	0.15	0.15
200 kHz < f ≤ 3 MHz	0.08	0.15
3 MHz < f ≤ 5 MHz	.15	0.38
5 MHz < f ≤ 30 MHz	0.30	0.38





Figure 1-3. Accuracy equations, K_i factor, and K_{osc} factor (test signal level $\leq 1 V_{rms}$)



$K_f \text{ Factor} : (\text{LONG}) \cdot (\text{MED,SHORT})$

Figure 1-4. Accuracy equations and K_f factor (test signal level $> 1 V_{rms}$)

Cable length factor

Add the following cable length factor A_c to the measurement accuracy when the cable length is set to 1 m (for 16048A) or 2 m (for 16048D) in *CABLE* field, after performing the cable correction and the OPEN/SHORT correction. When the cable length is 0 m, A_c is 0 percent.

$$A_c = \frac{f_m}{15} + A_{co} [\%]$$

A_{co} is the additional error when the impedance range is above 5 k Ω .

$$A_{co} = \frac{|Z_m| \cdot f_m \cdot K_t}{1000} [\%]$$

where:

- f_m = Test frequency in [MHz]
- Z_m = Absolute value of measured impedance in [k Ω]
- K_t = Test cable length in [m]

Temperature factor

Multiply the sum of the basic accuracy and the cable length factor by the following temperature induced error K_t , when the temperature range is 0 °C to 55 °C. The boundary belongs to the smaller multiplier.

Temperature (°C)	0	8	18	28	38	48	55
K_t		<3	<2	>1	<2	<3	>4

Figure 1-5. Temperature factor K_t

Measurement Accuracy Calculation Example

Example of L_s -Q accuracy calculation

Measurement conditions

Measured inductance L_x of DUT:	220 nH
Measured Q value of DUT:	30
Test signal level:	1 V _{rms}
Test frequency f_m :	25.2 MHz
Integration time:	LONG
Cable length:	0 m
Operating temperature:	28 °C

Determine inductance measurement accuracy A_e

1. From $|Z|$, $|Y|$, L , C , R , X , G , and B Accuracy (see page 6), measurement accuracy A_e is determined as below:

$$A_e = \pm(A_n + A_c) \times K_t$$

2. First of all, to determine the measurement accuracy A_e , calculate the impedance value from the DUT's inductance value. So the measurement impedance Z_m is:

$$Z_m = 2 \pi f_m L_x \\ \approx 35 [\Omega]$$

where:

- f_m = Test frequency [Hz]
- L_x = Measured inductance value of the DUT [H]

3. Choose an accuracy chart from Figure 1-3 and Figure 1-4. The oscillator level is 1 V_{rms}, then Figure 1-3 is chosen for this measurement.

4. Find the frequency point of f_m (25.2 MHz) along the X axis in Figure 1-3. Both axes are in log format. Interpolation may be required.

5. Find the impedance point of Z_m (35 Ω) along the Y axis in Figure 1-3 determined in step 2. Both axes are in log format. Interpolation may be required.

6. Mark the intersection of above two steps and determine the basic accuracy equation A_n , integration factor K_i , and oscillator level factor K_{osc} .

From:

Test frequency f_m :	25.2 MHz
DUT's impedance Z_m :	35 Ω
Integration time:	LONG
Test signal level:	1 V _{rms}

Then, $A_n = A_l$, $K_i = 1$, and $K_{osc} = 1$ (rounded from 0.02).

7. From Table 1-1, the actual accuracy equation to be applied is determined as A_I .

Then,

$$A_I = N_I \% + \left(\frac{f}{30}\right)^2 \cdot 3\% + \frac{50}{|Z_m|} [0.02\% + \left(\frac{f}{30}\right) \cdot 0.1\%] \cdot K_i \cdot K_{osc}$$

8. Determine N_I from Table 1-2.

From frequency = 25.2 MHz, then, $N_I = 0.3$.

9. Then,

$$A_n = 0.3\% + \left(\frac{25.2}{30}\right)^2 \cdot 3\% + \frac{50}{|35|} [0.02\% + \left(\frac{25.2}{30}\right) \cdot 0.1\%] \cdot 1 \cdot 1 \\ \approx 2.6 \text{ [%]}$$

10. Cable length is 0 m, then $A_c = 0$.

11. Operating temperature is 28 °C, then $K_t = 1$ (from Figure 1-5).

12. Therefore, inductance measurement accuracy A_e is:

$$A_e = \pm(A_n + A_c) \times K_t \\ = \pm(2.6 + 0) \times 1 \\ = \pm 2.6 \text{ [%]}$$

Determine Q measurement accuracy Q_e

1. From *Q Accuracy* (see page 6), Q measurement accuracy Q_e is determined as below:

$$Q_e = \pm \frac{Q_x^2 \times D_e}{1 \mp Q_x \times D_e}$$

2. Determine D accuracy D_e for calculating Q_e . From the previous step *Determine inductance measurement accuracy A_e* (see page 10), A_e is 2.6 [%], then:

$$D_e = \pm \frac{A_e}{100} \\ = \pm 0.026$$

3. Therefore Q_e is:

$$Q_e = \pm \frac{Q_x^2 \times D_e}{1 \mp Q_x \times D_e}$$

$$Q_e = \pm \frac{30^2 \times 0.026}{1 \mp 30 \times 0.026}$$

$$\approx +106/-13$$

Correction Functions

Zero open

Eliminates measurement errors due to parasitic stray admittance (C, G) of the test fixture.

Zero short

Eliminates measurement errors due to parasitic residual impedances (L, R) of the test fixture.

Load

Improves the measurement accuracy by using a device whose value is accurately known (a working standard) as a reference.

List Sweep

A maximum of ten frequencies or test signal levels can be programmed. Single or sequential tests can be performed. When Option 4285A-001 is installed, DC bias voltages can also be programmed.

Comparator Function

Ten-bin sorting for the primary measurement parameter, and IN/OUT decision output for the secondary measurement parameter.

Sorting modes

Sequential mode. Sorting into unnested bins with absolute upper and lower limits.

Tolerance mode. Sorting into nested bins with absolute or percent limits.

Bin count

0 to 999999

List sweep comparator

HIGH/IN/LOW decision output for each point in the list sweep table.

Other Functions

Store/Load

Ten instrument control settings, including comparator limits and list sweep programs, can be stored and loaded into and from the internal non-volatile memory. Ten additional settings can also be stored and loaded from each memory card.

GPIB

All control settings, measured values, comparator limits, and list sweep program can be controlled or monitored. Uses ASCII and 64-bit binary data format. GPIB buffer memory can store measured values for a maximum of 128 measurements and output packed data over the GPIB bus. Complies with IEEE-488.1 and 488.2. The programming language is Test and Measurement Systems Language (TMSL).

GPIB interface functions

SH1, AH1, T5, L4, SR1, RL1, DC1, DT1, C0, E1

Self test

Softkey controllable. Provides a means to confirm proper operation.

Option 4285A-001 (Internal DC Bias)

Adds the variable DC bias voltage function.

DC bias level

The following DC bias level accuracy is specified for an ambient temperature range of $23\text{ }^{\circ}\text{C} \pm 5\text{ }^{\circ}\text{C}$. Multiply the temperature induced setting error, K_t listed in Figure 1-5 for the temperature range of $0\text{ }^{\circ}\text{C}$ to $55\text{ }^{\circ}\text{C}$.

Voltage range	Resolution	Setting accuracy
$\pm(0.000\text{ to }4.000)\text{ V}$	1 mV	$\pm(0.1\% \text{ of setting} + 1\text{ mV})$
$\pm(4.002\text{ to }8.000)\text{ V}$	2 mV	$\pm(0.1\% \text{ of setting} + 2\text{ mV})$
$\pm(8.005\text{ to }20.000)\text{ V}$	5 mV	$\pm(0.1\% \text{ of setting} + 5\text{ mV})$
$\pm(20.01\text{ to }40.00)\text{ V}$	10 mV	$\pm(0.1\% \text{ of setting} + 10\text{ mV})$

A maximum DC bias current of 100 mA can be applied to the DUT.

DC bias monitor terminal

DC bias voltage or current can be monitored at the rear panel BNC connector.

The following monitor accuracies are applied when the digital volt meter whose input impedance is $\geq 10\text{ M}\Omega$ is used.

- DC bias voltage monitor

DC bias voltage across the DUT $\times 1$

Output impedance: 11 k Ω

Monitor accuracy: $\pm(0.2\% \text{ of reading} + 2 + 0.8 \times I_{\text{dut}})\text{ mV}$

where:

I_{dut} is current flowing through the DUT in [mA].

- DC bias current monitor

DC bias current through the DUT $\times 10\text{ }\Omega$ (1 V at 100 mA)

Output impedance: 10 k Ω

Monitor accuracy: $\pm(1\% \text{ of reading} + 0.3)\text{ mA}$

Other Options

Option 4285A-700: No DC bias

Option 4285A-002: Accessory Control Interface
Allows the 4285A to control the Agilent 42841A bias current source or the Agilent 42851A precision Q adapter.

The voltage ratio measurement accuracy, when the 4285A is used with the 42851A precision Q adapter, is described in the 42851A's operation manual.

Option 4285A-201: Handler interface

Option 4285A-202: Handler interface

Option 4285A-301: Scanner interface

Option 4285A-710: Blank panel

Option 4285A-907: Front handle kit

Option 4285A-908: Rack mount kit

Option 4285A-909: Rack flange and handle kit

Option 4285A-915: Add service manual

Option 4285A-ABA: Add English manual

Option 4285A-ABD: Add German manual

Option 4285A-ABJ: Add Japanese manual

Furnished Accessories

Power cord	Depends on the country where the 4285A is being used
100 Ω resistor box	Agilent P/N 04285-61001
BNC female-female Adapter	Agilent P/N 1250-0080 (4 ea.)
Fuse	Only for Option 4285A-201 Agilent P/N 2110-0046 (2 ea.)

Accessories Available

Test fixture/test leads

16034E	Test fixture for SMD or chip type DUT, $f \leq 40$ MHz
16047A	Test fixture for axial or radial DUT, $f \leq 13$ MHz
16047D	Test fixture for axial or radial DUT, $f \leq 40$ MHz
16048A	Test leads, length 1 m (BNC connector)
16048D	Test leads, length 2 m (BNC connector)
16065A	External voltage bias fixture
16334A	Tweezer-type test fixture for SMD or chip type DUT, $f \leq 15$ MHz
16451B	Dielectric test fixture
42842C	Bias current test fixture
42851-61100	SMD test fixture (Option 42842C-201)

DC bias source

42841A	Bias current source
--------	---------------------

Memory card

04278-89001	Memory card, 1 ea.
-------------	--------------------

GPIB interconnection cables

10833A	GPIB cable, 1 m
10833B	GPIB cable, 2 m
10833C	GPIB cable, 4 m
10833D	GPIB cable, 0.5 m

Power Requirements

Line voltage

100, 120, 220 Vac $\pm 10\%$, 240 Vac $+5\% -10\%$

Line frequency

47 to 66 Hz

Power consumption

200 VA max.

Operating Environment

Temperature

0 °C to 55 °C

Humidity

$\leq 95\%$ R.H. at 40 °C

Dimensions

426 (W) by 177 (H) by 498 (D) (mm)

Weight

Approximately 16 kg (35.3 lb., standard)

Display

LCD dot-matrix display

Capable of displaying

Measured values
Control settings
Comparator limits and decisions
List sweep tables
Self test message and annunciations

Number of display digits

6-digits, maximum display count 999999

Supplemental Performance Characteristics

The Agilent 4285A supplemental performance characteristics are listed below. These supplemental performance characteristics are not specifications, but are typical characteristics included as supplemental information for the operator.

Stability

When the following conditions are satisfied,

Integration time: LONG
 Operating temperature: Constant operating temperature of 23 °C ± 5 °C

Parameter	≤ 1 MHz	30 MHz
Z , Y , L, C, R	< 0.01%/day	< 0.05%/day
D	< 0.0001/day	< 0.0005/day

Temperature coefficient

When the following conditions are satisfied,

Integration time: LONG
 Test signal voltage: ≥ 20 mV_{rms}
 Operating temperature: 23 °C ± 5 °C

Parameter	≤ 1 MHz	30 MHz
Z , Y , L, C, R	< 0.004%/°C	< 0.05%/°C
D	< 0.00004/°C	< 0.0005/°C

Settling time

Frequency (f_m)

< 50 msec.

Test signal level

< 100 msec.

Measurement range

< 50 msec./range shift

Input protection

Internal circuit protection, when a charged capacitor is connected to the UNKNOWN terminals.

The maximum capacitor voltage is:

$$V_{max} = \sqrt{\frac{1}{C}} [V]$$

where:

$V_{max} = \leq 200$ V
 C = Capacitance value in Farads

Measurement time

Typical measurement times from the trigger to the output of EOM at the Handler Interface. (EOM: End of Measurement)

Integration time	Measurement time
SHORT	30 ms
MEDIUM	65 ms
LONG	200 ms

In the following condition an additional measurement time, approx. 300 ms, is added to the measurement time.

Test signal voltage: 0.51 V – 2 V_{rms}
 Measurement range: 0 Ω Range
 Test signal current: ≥ 22 mA

Display time

Display time for each display format is given as:

MEAS DISPLAY page Approx. 8 ms
BIN No. DISPLAY page Approx. 5 ms
BIN COUNT DISPLAY page Approx. 0.5 ms

GPIB data output time

Internal GPIB data, processing time from EOM output to measurement data output on GPIB lines (excluding display time):

- Approx. 10 ms

Option 4285A-001 (internal DC bias)

Maximum DC bias current when the normal measurement can be performed is 100 mA.

DC bias settling time

When DC bias is set to ON, add 5 ms to the measurement time. This settling time does not include the DUT charge time.

Sum of DC bias settling time plus DUT (capacitor) charge time is shown in the following figure.




Figure 1-6. Sum of the DC bias settling time and DUT (capacitor) charge time

Web Resource

www.agilent.com/find/lcrmeters

Remove all doubt

Our repair and calibration services will get your equipment back to you, performing like new, when promised. You will get full value out of your Agilent equipment throughout its lifetime. Your equipment will be serviced by Agilent-trained technicians using the latest factory calibration procedures, automated repair diagnostics and genuine parts. You will always have the utmost confidence in your measurements. Agilent offers a wide range of additional expert test and measurement services for your equipment, including initial start-up assistance onsite education and training, as well as design, system integration, and project management.

For more information on repair and calibration services, go to:

www.agilent.com/find/removealldoubt



Agilent Email Updates

www.agilent.com/find/emailupdates

Get the latest information on the products and applications you select.



Agilent Direct

www.agilent.com/find/agilentdirect

Quickly choose and use your test equipment solutions with confidence.

www.agilent.com

For more information on Agilent Technologies' products, applications or services, please contact your local Agilent office. The complete list is available at:

www.agilent.com/find/contactus

Americas

Canada	(877) 894-4414
Latin America	305 269 7500
United States	(800) 829-4444

Asia Pacific

Australia	1 800 629 485
China	800 810 0189
Hong Kong	800 938 693
India	1 800 112 929
Japan	0120 (421) 345
Korea	080 769 0800
Malaysia	1 800 888 848
Singapore	1 800 375 8100
Taiwan	0800 047 866
Thailand	1 800 226 008

Europe & Middle East

Austria	0820 87 44 11
Belgium	32 (0) 2 404 93 40
Denmark	45 70 13 15 15
Finland	358 (0) 10 855 2100
France	0825 010 700*
	*0.125 €/minute
Germany	01805 24 6333**
	**0.14 €/minute
Ireland	1890 924 204
Israel	972-3-9288-504/544
Italy	39 02 92 60 8 484
Netherlands	31 (0)20 547 2111
Spain	34 (91)631 3300
Sweden	0200-88 22 55
Switzerland	0800 80 53 53
United Kingdom	44 (0) 118 9276201

Other European Countries:

www.agilent.com/find/contactus

Revised: March 27, 2008

Product specifications and descriptions in this document subject to change without notice.

© Agilent Technologies, Inc. 1995, 2000, 2003, 2008

Printed in USA, April 14, 2008

5963-5395E



Agilent Technologies