Data sheet

100 MHz Analog Oscilloscope

Model 2190B

- Dual time base oscilloscope (2 channel)
- 5 mV/division sensitivity
- Sweeps to 5 ns/division
- 23 calibrated ranges, main time base
- Signal delay line
- 12 kV accelerating voltage
- Channel 2 output



Specifications	mode	
	2190B	
ERTICAL AMPLIFIERS (CI	H I and CH 2)	
Sensitivity	5 mV/div to 5 V/div. 1 mV/div to 1V/div (at X5 MAC	
Attenuator	10 calibrated steps in 1-2-5 sequence.	
	Vernier control provides fully adjustable sensitivity	
	between steps, adjustment range 1/1 to 1/3	
Accuracy	±3% (±5% at X5 MAG)	
Input Impedance	I MΩ +2%	
Input Capacitance	25 pF ±10 pF	
Frequency Response	DC: DC to 100 MHz (-3 dB)	
X5 MAG	DC to 25 MHz (-3 dB)	
AC	10 Hz to 100 MHz (-3 dB)	
Rise Time	3.5 ns (Overshoot ≤5%)	
Signal Delay Time	Variable	
Square Wave Characteristics	Overshoot less than 5%, 10 mV/div range	
Square wave characteristics	Other ranges within 5% additional	
Maximum Input Voltage	400 V (DC + AC peak)	
Maximum Input Voltage	1 400 V (DC + AC peak)	
ERTICAL AMPLIFIERS		
Operating Modes	CH 1, CH 2, Dual, Add	
Delay Time Between Channels	Within I ns between CH I and CH 2	
Crosstalk	30:1 at 100 kHz	
WEED CYCTEM		
WEEP SYSTEM		
Operating Modes		
A	A sweep	
В	Delayed B sweep	
B TRIGGERED	B sweep triggered after delay	
A Time Base		
Sweep Mode	Main, Mix, Delay, XY	
Sweep Time:	5 s to 20 ns/div., 23 steps in 1-2-5 sequence	
	with variable control	
Accuracy	± 3%	
Hold Off Time	Continuously variable. Adjustment range from	
	normal to 5 times normal	
B Time Base		
Delay Method	Continuous delay. Triggered delay	
Sweep Time	20 ns. to 0.5 s/div., 23 steps in 1-2-5 sequence	
Accuracy	± 3%	
Delay Time	Start point: 0.5 div to $+ 0.3$ div.	
Belay Time	End point: 10 div + 1 div	
Delay Jitter	Within 1/10,000 of full scale sweep time	
Delay Jillei	Within 1/10,000 of full scale sweep time	
RIGGERING		
A Trigger		
Source	CH 1, CH 2, LINE, EXT, ALT	
Sensitivity	30 Hz to 110 MHz	
	1.5 div (internal), ≥0.5 p-p (external)	
TV-V	20 Hz - 30 kHz	
TV-H	1.0 div (internal), ≥0.5 p-p (external) 3 kHz - 100 kHz	
	1.0 div (internal), ≥0.5 p-p (external)	
Slope	+ or -	
B Trigger	The A trigger is also the B trigger	

Mandanian Innest Maltana	200 V (DC + AC
Maximum Input Voltage	300 V (DC + AC peak)
	_
HORIZONTAL AMPLIFIE	R
X-Y Mode	X Axis = CH I. Y Axis = CH 2
Sensitivity	5 mV/div to 5 V/div, CH I and CH 2
Accuracy	±3% calibrated position, ±6% using x10 MAG
Frequency Response	DC to 2 MHz (-3 dB)
CH2 (Y) OUTPUT	
Output Voltage	Approx. 100 mV/div open circuit
	Approx. 50 mV/div into 50 Ω
Freq. Response	20 Hz to 100 MHz, -3 db
Output Impedance	approx. 50 Ω
CRT	
Туре	Rectangular with integral graticule
Display Area	8 x 10 div (1 div = 1 cm)
Accelerating Voltage	12 kV
Phosphor	P31
Scale Illumination	None
	Electrical, front panel adjustable

Other Specifications		
Z Axis	Sensitivity: 3 V or greater, TTL level.	
(Intensity Modulation)	Intensity increasing with more positive levels	
Input Impedance	50 kΩ	
Usable Freq. Range	DC to 5 MHz	
Maximum Input Voltage	30 V (DC + AC peak)	
CAL/Probe Compensation		
Waveform	Positive going squareware	
Output Voltage	2 V p-p ±3%	
Frequency	Approx. I kHz	
Power Requirements	$100/120/220/240/$ VAC $\pm 10\%$, $50/60$ Hz,	
	approximately 55 W	
Dimensions (HxWxD)	12.76 x 15.68 x 5.2" (324 x 398 x 132 mm)	
Weight	18.7 lbs (8.5 kg)	
ENVIRONMENT		
Within Specified Accuracy	50° to 95°F (10° to 35°C), 10-80% RH	
Full Operation	32° to 122°F (0° to +50°C), 10-80% RH	
Storage	-22° to 158°F (-30° to +70°C), 10-90% RH	
	Three Year Warranty	

Accessories

Supplied: Instruction Manual, Two PR 37A x1/x10/Ref. Probes or equivalent,

AC Power Cord, Spare Fuse

Optional: PR 32A Demodulator Probe, PR 37AG x1/x10/REF. Probe,

PR 100A x100 Probe, PR-55 High Voltage x1000 Probe,

LC 210A Carrying Case

