
**SPANAWAVE** 

CK519A: 4-in-1 OSLT Calibration Kit, DC to 18 GHz, Type-N (f) 50 Ohm



## Interface

According to

Type-N (f)

## **Contents and Documentation**

This kit is delivered with

- **Standard Definitions Card** •
  - Printed Standard Definitions that can be used on nearly all Vector Network Analyzers
- **Test Results Documentation** •
- Lanyard
- Hard Shell Case

### Material and plating

**Connector parts** Center conductor Outer conductor Body Dielectric Substrate

Material Stainless steel Aluminum PPE Al<sub>2</sub>O<sub>3</sub>

Plating Beryllium copper Gold, min. 1.27 µm, over nickel Passivated black anodized

\*Prices are for US customers only. International prices may differ based on region.

Spanawave Corp.   Tel: 866-202-9262   www.spanawave.com	Page 1 / 3



# CK519A: 4-in-1 OSLT Calibration Kit, DC to 18 GHz, Type-N (f) 50 Ohm

Electrical data	
Frequency range	DC to 18 GHz
Thru	
Return loss	<ul> <li>≥ 40 dB, DC to 6 GHz</li> <li>≥ 36 dB, 6 GHz to 9 GHz</li> <li>≥ 32 dB, 9 GHz to 18 GHz</li> </ul>
<u>Open</u>	
Error from nominal phase <sup>1</sup>	$\leq$ 2.0°, DC to 6 GHz
	$\leq$ 3.0°, 6 GHz to 9 GHz
	$\leq$ 4.0°, 9 GHz to 18 GHz
<u>Short</u>	
Error from nominal phase <sup>2</sup>	$\leq$ 1.5°, DC to 6 GHz
	$\leq$ 2.0°, 6 GHz to 9 GHz
	$\leq$ 2.5°, 9 GHz to 18 GHz
Load	
Return loss	$\geq$ 42 dB, DC to 6 GHz
	$\geq$ 36 dB, 6 GHz to 9 GHz
	$\geq$ 30 dB, 9 GHz to 18 GHz
DC-Resistance	$50 \Omega \pm 0.5 \Omega$
Power handling	$\leq$ 1.0 W

<sup>1</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Fringing Capacitances.

 $^2$  The nominal phase is defined by the Offset Delay, the Offset Loss and the Short Inductance.

# Mechanical data

Rev 042518

Mating cycles	≥ 500
Maximum torque	1.70 Nm
Recommended torque	1.10 Nm
Gauge	5.22 mm to 5.26 mm

### General standard definitions

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

<u>Thru</u> Offset Z <sub>o</sub> / Impedance / Z <sub>o</sub> Offset Delay Length (electrical) / Offset Length Offset Loss Loss Line Loss @ 1GHz	50 Ω 152.105 ps 45.60 mm 2.20 GΩ/s 0.0291 dB/ √GHz 0.0006 dB/mm	
<b>Open</b> Offset Z <sub>o</sub> / Impedance / Z <sub>o</sub> Offset Delay Length (electrical) / Offset Length Offset Loss Loss Fringing Capacitances	50 Ω 40.028 ps 12.00 mm 0.80 GΩ/s 0.0056 dB/ $\sqrt{GHz}$ C <sub>0</sub> = 37.1000 x 10 <sup>-15</sup> F /	37.1000 fF
	$C_1 = 1200.00 \times 10^{-27} \text{ F/Hz} /$	1.20000 fF /GHz
	$C_2 = -30.0000 \times 10^{-36} \text{ F/Hz}^2$ /	-0.03000 fF /GHz <sup>2</sup>
	$C_3 = 0.00000 \times 10^{-45} \text{ F/Hz}^3 /$	0.00000 fF /GHz <sup>3</sup>

Spanawave Corp.   Tel: 866-202-9262   www.spanawave.com	Page 2 / 3

# **Technical Data Sheet**

**SPANAWAVE** 

CK519A: 4-in-1 OSLT Calibration Kit, DC to 18 GHz, Type-N (f) 50 Ohm

#### <u>Short</u>

Offset Z <sub>o</sub> / Impedance / Z <sub>o</sub>	50 Ω	
Offset Delay	40.028 ps	
Length (electrical) / Offset Length	12.00 mm	
Offset Loss	0.80 GΩ/s	
Loss	0.0056 dB/ √GHz	
Short Inductance	$L_0 = 95.0000 \times 10^{-12} H$ /	95.0000 pH
	$L_1 = -9900.00 \times 10^{-24} \text{ H/Hz} /$	-9.90000 pH/GHz
	$L_2 = 980.000 \times 10^{-33} \text{ H/Hz}^2 /$	0.98000 pH/GHz <sup>2</sup>
	$L_3 = -29.0000 \times 10^{-42} \text{ H/Hz}^3 /$	-0.02900 pH/GHz <sup>3</sup>
Load		
Offset $Z_o$ / Impedance / $Z_o$	50 Ω	
Offset Delay	0.0000 ps	
Length (electrical) / Offset Length	0.000 mm	
Offset Loss	0.00 GΩ/s	
Loss	0.0000 dB/ √GHz	

Environmental data	
Operating temperature range <sup>3</sup>	+20 °C to +26 °C
Rated temperature range of use <sup>4</sup>	0 °C to +50 °C
Storage temperature range	- 40 °C to +85 °C
RoHS	compliant

<sup>3</sup> Temperature range over which these specifications are valid.

<sup>4</sup> This range is underneath and above the operating temperature range, within the calibration kit is fully functional and could be used without damage.

#### Includes

Standard delivery for this kit includes Test Results. The documentation issued reports which quantities were tested individually, traceable to national / international standards. Model based standard definitions of the calibration standards are reported in Agilent / Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Calibration interval Recommendation	12 months
Packing Standard Weight	1 per bag 8.3 oz.

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.